




## OPERATOR MANUAL ADDENDUM

### Environics Model 6113

The Environics 6113 is a variation of the Environics 6103, with only a single Mass Flow Controller (MFC). A number of changes have been made to the Environics 6103 software, which differ from what is described in the 6103 User Manual

- Gases Setup (Pg 15) has been removed. Since the system only has a single MFC, gas mixing is not possible, so the GASES setup menu has been removed.
- FLOW MODE (pg 24) has been removed. Since only a single MFC is used, the same functionality is provided by CONC mode.
- **IMPORTANT:** Do not attempt to use the automated leak test. In its final version, this mode will be functional.
- When connecting an analyzer, that has an unvented input, make sure the vent port of the 6113 is not capped.



**Envirotronics Inc.**

TOLSON, IL 60084

TITLE

S6113 TRANSFER STANDARD

REV. DATE REVISION NO. MANUFACTURE NUMBER

A

6113 TRANSFER

SCALE: 1" = 1" CUSTOMER: B 5/76

JUN 1 1976